

# Rat IgG1 kappa Isotype Control (eBRG1), PerCP-eFluor™ 710, eBioscience™

| Product Details         |                                     |
|-------------------------|-------------------------------------|
| Size                    | 25 µg                               |
| Host/Isotype            | Rat / IgG1, kappa                   |
| Class                   | Control                             |
| Type                    | Isotype Control                     |
| Clone                   | eBRG1                               |
| Conjugate               | PerCP-eFluor™ 710                   |
| Excitation/Emission Max | 482/708 nm                          |
| Form                    | Liquid                              |
| Concentration           | 0.2 mg/mL                           |
| Purification            | Affinity chromatography             |
| Storage buffer          | PBS, pH 7.2                         |
| Contains                | 0.09% sodium azide                  |
| Storage conditions      | 4° C, store in dark, DO NOT FREEZE! |
| RRID                    | AB_1944422                          |

| Applications          | Tested Dilution | Publications  |
|-----------------------|-----------------|---------------|
| Flow Cytometry (Flow) | Assay-Dependent | 0 Publication |
| Control (Ctrl)        | Assay-Dependent | -             |

## Product Specific Information

**Description:** The monoclonal rat IgG1, kappa is useful as an isotype control immunoglobulin.

**Applications Reported:** PerCP-eFluor® 710 Rat IgG1 Isotype Control has been reported for use in flow cytometric analysis.

**Applications Tested:** PerCP-eFluor® 710 Rat IgG1 Isotype Control has been tested by flow cytometric analysis of mouse splenocytes and normal human peripheral blood cells. Isotype control should be used at the same concentration as experimental antibody.

PerCP-eFluor® 710 can be used in place of PE-Cy5, PE-Cy5.5 or PerCP-Cy5.5. PerCP-eFluor® 710 emits at 710 nm and is excited with the blue laser (488 nm). Please make sure that your instrument is capable of detecting this fluorochrome. For a filter configuration, we recommend using the 685 LP dichroic mirror and 710/40 band pass filter, however the 695/40 band pass filter is an acceptable alternative.

Our testing indicates that PerCP-eFluor® 710 conjugated antibodies are stable when stained samples are exposed to freshly prepared 2% formaldehyde overnight at 4°C, but please evaluate for alternative fixation protocols.

**Excitation:** 488 nm; **Emission:** 710 nm; **Laser:** Blue Laser.

**Filtration:** 0.2 µm post-manufacturing filtered.

## 2 References

VprBP Is Required for Efficient Editing and Selection of Ig+ B Cells, but Is Dispensable for Ig+ and Marginal Zone B Cell Maturation and Selection. J Immunol (2015)

Chronic thoracic spinal cord injury impairs CD8+ T-cell function by up-regulating programmed cell death-1 expression. J Neuroinflammation (2014)

For Research Use Only. Not for use in diagnostic procedures. Not for resale without express authorization. Products are warranted to operate or perform substantially in conformance with published Product specifications in effect at the time of sale, as set forth in the Production documentation, specifications and/or accompanying package inserts ("Documentation"). No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is limited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than the Buyer. Any model or sample furnished to Buyer is merely illustrative of the general type and quality of goods and does not represent that any Product will conform to such model or sample. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE GRANTED INCLUDING WITHOUT LIMITATION, IMPLIED WARRANTIES OF MERCHANTABILITY, FITNESS FOR ANY PARTICULAR PURPOSE, OR NON INFRINGEMENT. BUYER'S EXCLUSIVE REMEDY FOR NON-CONFORMING PRODUCTS DURING THE WARRANTY PERIOD IS LIMITED TO REPAIR, REPLACEMENT OF OR REFUND FOR THE NON-CONFORMING PRODUCT(S) AT SELLER'S SOLE OPTION. THERE IS NO OBLIGATION TO REPAIR, REPLACE OR REFUND FOR PRODUCTS AS THE RESULT OF (I) ACCIDENT, DISASTER OR EVENT OF FORCE MAJEURE, (II) MISUSE, FAULT OR NEGLIGENCE OF OR BY BUYER, (III) USE OF THE PRODUCTS IN A MANNER FOR WHICH THEY WERE NOT DESIGNED, OR (IV) IMPROPER STORAGE AND HANDLING OF THE PRODUCTS. Unless otherwise expressly stated on the Product or in the documentation accompanying the Product, the Product is intended for research only and is not to be used for any other purpose, including without limitation, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses, or any type of consumption by or application to human or animals.